



Ref. Certif. No.

SI-3279

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST  
CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE)  
CB SCHEMESYSTEME CEI D'ACCEPTATION MUTUELLE DE  
CERTIFICATS D'ESSAIS DES EQUIPEMENTS  
ELECTRIQUES (IECEE) METHODE OC

## CB TEST CERTIFICATE

## CERTIFICAT D'ESSAI OC

Product  
Produit

Power Supply for Medical use

Name and address of the applicant  
Nom et adresse du demandeurFRIWO Gerätebau GmbH  
Von-Liebig Strasse 11, D-48346 Ostbevern, GermanyName and address of the manufacturer  
Nom et adresse du fabricantFRIWO Gerätebau GmbH  
Von-Liebig Strasse 11, D-48346 Ostbevern, GermanyName and address of the factory  
Nom et adresse de l'usine1) FRIWO Gerätebau GmbH  
Von-Liebig Strasse 11, D-48346 Ostbevern, GermanyNote: When more than one factory, please report on page 2  
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2<sup>ème</sup> page Additional Information on page 2Ratings and principal characteristics  
Valeurs nominales et caractéristiques principalesInput: 100-240 Va.c.; 50-60 Hz; 2,0-0,7 A; Class I  
Output:  Additional Information on page 2Trademark (if any)  
Marque de fabrique (si elle existe)

FWGB

Type of Manufacturer's Testing Laboratories used  
Type de programme du laboratoire d'essais constructeur

/

Model / Type Ref.  
Ref. De typeType: DT150-24 and DT150-28  
 Additional Information on page 2Additional information (if necessary may also be reported  
on page 2)  
Les informations complémentaires (si nécessaire,, peuvent  
être indiqués sur la 2<sup>ème</sup> pageThe risk management requirements of the standard were not  
addressed.  
 Additional Information on page 2A sample of the product was tested and found  
to be in conformity with  
Un échantillon de ce produit a été essayé et a été  
considéré conforme à la

IEC 60601-1:2005 (3rd Ed.)

As shown in the Test Report Ref. No. which forms part of  
this Certificate  
Comme indiqué dans le Rapport d'essais numéro de  
référence qui constitue partie de ce Certificat

T223-0126/12, dated 2012-04-21

This CB Test Certificate is issued by the National Certification Body  
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**Slovenski institut za kakovost in meroslovje  
Slovenian Institute of Quality and Metrology  
Tržaška c. 2, SI-1000 Ljubljana, Slovenia  
Product Certification Body is accredited by Slovenian Accreditation, Reg. No.: CP-001

Date: 2012-04-23

Signature: Alja Pregl



**Name and address of the factory**

1) FRIWO Gerätebau GmbH

Von-Liebig Strasse 11, D-48346 Ostbevern, Germany

2) TABEMAX SP J

ULICA BETONOWA 1

86-005 BIALE BLOTA, POLAND

3) TABER

Ul. Oliwiana 24

85-461 Bydgoszcz, Poland

4) NOVIL Sp. z.o.o.

Miszewko, Sloneczna 8

80-297 Banino, Poland

5) Golden Regent Electronics Industrial Ltd.

Huangjinghu Management District, 523500 Qishi Town,

Dongguan, Guangdong, PR China

6) Roofer Technology (Shenzehn) Co., Ltd.

Roofer Building, GeXia Industrial Area,

Sanlian Village, Dalang Road, Longhua Town, Shenzhen, PR China

**Ratings and principal characteristics:**

Input: 100-240 Vac; 50-60 Hz; 2,0-0,7 A; Class I

Output of the power supply type: DT150-24: 24 Vdc; 6,25 A

Output of the power supply type: DT150-28: 28 Vdc; 5,36 A

**Additional information (if necessary)****Information complémentaire (si nécessaire)**

Date: 2012-04-23

Signature: Alja Pregl

